



PUBLIC NOTICE

Intertek Warns of Fraudulent IECEE CB Scheme Test Certificate US/5096/ITS for RFVII Inc. Model RF-3-III RF Generator

Published: 13 October 2016

The following is a notification from Intertek that the RFVII Inc. Model **RF-3-III** RF Generator IECEE CB Scheme Test Certificate **US/5096/ITS** is fraudulent. The RFVII Inc. Model RF-3-III has not been evaluated or tested by Intertek and it is not known if the product complies with any safety requirements.

Product Description	<i>RFVII Inc. Model RF-3-III RF Generator, rated 100-240 VAC, 3 A, 50/60 Hz</i>
Manufacturer	<i>RFVII, Inc.</i>
Reported Hazards	<i>Unknown</i>
What to Do	<p><i>If the fraudulent test certificate is or has been presented to show compliance, it should be immediately rejected and reported to Intertek and the IECEE Secretariat at: secretariat@iecee.org</i></p> <p><i>RFVII Inc. has indicated that the certificate distribution was limited and all copies have been recalled and destroyed. Please note, this notice does not include IECEE Test Certificate ITS/5096/ITS covering Model RF-1-III which is valid</i></p>

		Ref. Certif. No. US/5096/ITS
IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME		SYSTEM CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC
CB TEST CERTIFICATE		CERTIFICAT D'ESSAI OC
Product	RF Generator	
Name and address of the Applicant	RFVII, Inc. 1041 Glassboro Rd, Building EKL, Williams Town, NJ 08094 United States of America	
Name and address of the manufacturer	RFVII, Inc. 1041 Glassboro Rd, Building EKL, Williams Town, NJ 08094 United States of America	
Name and address of the factory	RFVII, Inc. 1041 Glassboro Rd, Building EKL, Williams Town, NJ 08094 United States of America	
Rating and principal characteristics	100 - 240VAC, 3A, 50/60Hz	
Trademark (if any)	RFVII	
Type of manufacturer's Testing Laboratories used	RF-3-III	
Model / Type Ref.	RF-3-III	
Additional information (if necessary may also be reported on page 2)	61010-1(ed.3)	
A sample of product was tested and found to be in conformity with IEC	CA, CH, JP, US	
National differences / Comments	101502008MPK-001	
This CB Test Certificate is issued by the National Certification Body: Intertek Testing Services, N.A. 165 Main Street, Cortland, NY 13045, USA		
Ce Certificat d'essai OC est établi par l'Organisme National de Certification 		
Date: 2014-05-14	Signature: John Quigley	page 1 of 1

FRAUDULENT TEST CERTIFICATE